

**Foshan Flying Medical Products Co., Ltd**  
**CERTIFICATE OF STERILIZATION ANALYSIS**

Manufacturer		Sleek Brows Ltd., for Everlasting Brows		Product Description		EVERLASTING BROWS DISPOSABLE TOOL	
Country of Manufactory		P.R.China		Expiration Date		25/08/2022	
Size		5*20		Lot Number		EVER030	
Item#		EL01		Inspection Date		29/04/2018	
PO#		PO180323Z21		Customer Name		Everlasting Brows	
Flying PO#		PN180323Z21		QTY		13000pcs	
NO	Test Item	Requirements	Inspection Standard	Inspected Qty.	Result	Conclusion	
1	Appearance	Clean, level off, no besmirch, no breakage	S-3	13000pcs	Accord	Pass	
2	Package Size	5CM*20CM Dialysis bag	S-2	13000pcs	Accord	Pass	
3	Box size	40cm*40cm*40cm			Accord	Pass	
4	Shear Strength	≥1N/cm			Accord	Pass	
5	Unwrap Quality	Good and ease to unwrap			Accord	Pass	
6	Seal	Sealed			Accord	Pass	
7	Printing Quality	Good quality good align				Pass	
8	Sterilization	Sterility	N/A	13000pcs	Accord	Pass	
Final Conclusion			Inspection Passed				

**Drafted by** Jane Zhang  
**Date** :29/04/2018



**Approved by** Jane Zhang  
**Date** :29/04/2018

Foshan Flying Medical Products Co., Ltd

Sterilization Inspection Report

NO :20180510003

PO#	PO180323Z21		Product	EVERLASTING BROWS DISPOSABLE TOOL			
LOT	L1061010		Item#	EL01			
Sterilization Date	03/05/2018		Size	5*20			
Sterilization Lot	EVER030		Amount	13000pcs			
Process Parameter Summary			Inspection Basis	EN ISO 11737			
Temperature	50 ±2		Humidity	45-80%			
Exposure Time(h)	10h		EO Weight(kg)	105 kgs			
Biological Indicator Status							
Manufacture	Hangzhou Future		Organism	ATCC 9372			
Lot	20180303		D-value	3.5min			
Population	2.2*10 C.F.U		Test Quantity	10+1 strips			
Incubation Situation							
Incubation Date	2018.5.3-2018.5.10		Incubation Temperature	35±2			
Incubation Time	168		Medium Name	Soybean Casein Digest Medium			
Incubation Time	After 24h	After 48h	After 72h	After 96h	After 120h	After 144h	After 168h
Observed Item	(-)	(-)	(-)	(-)	(-)	(-)	(-)
Positive Control	(+)						
<p>Conclusion Qualified</p> <p>Qualified</p> <p>Sterilized by EO,the products comply with ISO 11737 standard(Sterilization of medical devices--Microbiological methods--Part2:Tests of sterility performed in the validation of a sterilization process).</p>							

Note

(-) represent of negative BI,no bacterial growth

(+) represent of negative BI,bacterial growth

Drafted by Jane Zhang

Approved by Jane Zhang

Date:10/05/2018

Date : 10/05/2018



A red circular stamp with the text "EVERLASTING MEDICAL PRODUCTS" around the perimeter. In the center, there is Chinese text: "佛山恒美医药有限公司" and "FOSHAN HENGMEI MEDICAL PRODUCTS CO., LTD.". A signature in black ink, which appears to be "Jane Zhang", is written across the stamp.