

Foshan Flying Medical Products Co., Ltd

CERTIFICATE OF STERILIZATION ANALYSIS

Manufacturer		Sleek Brows Ltd., for Everlasting Brows		Product Description		EVERLASTING BROWS HIGH QUALITY STAINLESS-STEEL MICROBLADING BLADES	
Country of Manufactory		P.R.China		Expiration Date		29/12/2027	
Size		10F.18/12F/18/14F.18 10F.25/18U.15/18U.18		Lot Number		22122401	
Item#		14U,0.25mm		Inspection Date		31/12/2022	
PO#		221201Z21		Customer Name		Everlasting Brows	
Flying PO#		PN221201Z21		QTY		3000boxes	
NO	Test Item	Requirements	Inspection Standard	Inspected Qty.	Result	Conclusion	
1	Appearance	Clean, level off, no besmirch, no breakage	S-3	3000boxes	Accord	Pass	
2	Package Size	20pcs/box	S-2	3000boxes	Accord	Pass	
3	Box size	20mm*40mm*100mm			Accord	Pass	
4	Shear Strength	≥1N/cm			Accord	Pass	
5	Unwrap Quality	Good and ease to unwrap			Accord	Pass	
6	Seal	Sealed			Accord	Pass	
7	Printing Quality	Good quality good align				Pass	
8	Sterilization	Sterility	N/A	3000boxes	Accord	Pass	
Final Conclusion			Inspection Passed				

Drafted by Jane Zhang

FOSHAN FLYING MEDICAL PRODUCTS
佛山飞鹰医疗器械有限公司
Jane Zhang

Approved by Jane Zhang

Date :2022.12.31

Date :2022.12.31

Foshan Flying Medical Products Co., Ltd

Sterilization Inspection Report

NO

PO#	PO221201Z21		Product	EVERLASTING BROWS HIGH QUALITY STAINLESS-STEEL MICROBLADING BLADES			
LOT	22122401		Item#	10F.18/12F.18/14F.18 10F.25/18U.15/18U.18			
Sterilization Date	24/12/2022		Size	MIXED			
Sterilization Lot	22122425		Amount	3000boxes			
Process Parameter Summary			Inspection Basis	EN ISO 11737			
Temperature	50±2		Humidity	45-80%			
Exposure Time(h)	14h		EO Weight(kg)	60 kgs			
Biological Indicator Status							
Manufacture	Hangzhou Future		Organism	ATCC 9372			
Lot	221910		D-value	3.5min			
Population	2.2*10 C.F.U		Test Quantity	10+1 strips			
Incubation Situation							
Incubation Date	2022-12-24- 2022-12-31		Incubation Temperature	35±2			
Incubation Time	168		Medium Name	Soybean Casein Digest Medium			
Incubation Time	After 24h	After 48h	After 72h	After 96h	After120h	After144h	After168h
Observed Item	(-)	(-)	(-)	(-)	(-)	(-)	(-)
Positive Control	(+)						
Conclusion	: Qualified Qualified						

Sterilized by EO,the products comply with ISO 11737 standard(Sterilization of medical devices-- Microbiological methods--Part2:Tests of sterility performed in the validation of a sterilization process).

Note

- (-) represent of negative BI,no bacterial growth
- (+) represent of negative BI,bacterial growth

Drafted by Jane Zhang

Date :2022.12.31

Approved by Jane Zhang

Date : 2022.12.31



The image shows a handwritten signature in black ink that reads "Jane Zhang". The signature is written over a red circular stamp. The stamp contains the text "EVERLASTING MEDICAL PRODUCTS" around the top edge and Chinese characters in the center. There is also a small star symbol at the bottom of the stamp.