

Foshan Flying Medical Products Co., Ltd

CERTIFICATE OF STERILIZATION ANALYSIS

Manufacturer	Sleek Brows Ltd., for Everlasting Brows		Product Description	EVERLASTING BROWS HIGH QUALITY STAINLESS-STEEL MICROBLADING BLADES		
Country of Manufactory	P.R.China		Expiration Date	01/09/2026		
Size	14U.25		Lot Number	210831		
Item#	14U,0.25mm		Inspection Date	31/09/2021		
PO#	210729Z21		Customer Name	Everlasting Brows		
Flying PO#	PN210729Z21		QTY	400boxes		
NO	Test Item	Requirements	Inspection Standard	Inspected Qty.	Result	Conclusion
1	Appearance	Clean, level off, no besmirch, no breakage	S-3	400boxes	Accord	Pass
2	Package Size	20pcs/box	S-2	400boxes	Accord	Pass
3	Box size	20mm*40mm*100mm			Accord	Pass
4	Shear Strength	≥1N/cm			Accord	Pass
5	Unwrap Quality	Good and ease to unwrap			Accord	Pass
6	Seal	Sealed			Accord	Pass
7	Printing Quality	Good quality good align				Pass
8	Sterilization	Sterility	N/A	400boxes	Accord	Pass
Final Conclusion		Inspection Passed				

Drafted by Jane Zhang



Approved by Jane Zhang

Date :2021.8.31

Date :2021.8.31

Foshan Flying Medical Products Co., Ltd

Sterilization Inspection Report

NO

PO#	PO210729Z21	Product	EVERLASTING BROWS HIGH QUALITY STAINLESS-STEEL MICROBLADING BLADES				
LOT	210831	Item#	14U.25				
Sterilization Date	010/9/2021	Size	14U,0.25mm				
Sterilization Lot	21191025	Amount	400boxes				
Process Parameter Summary		Inspection Basis	EN ISO 11737				
Temperature	50±2	Humidity	45-80%				
Exposure Time(h)	10h	EO Weight(kg)	7.5kgs				
Biologocal Indicator Status							
Manufacture	Hangzhou Future	Organism	ATCC 9372				
Lot	21191025	D-value	3.5min				
Population	2.2*10 C.F.U	Test Quantity	10+1 strips				
Incubation Situation							
Incubation Date	2021-9-1- 2021-9-8		Incubation Temperature	35±2			
Incubation Time	168		Medium Name	Soybean Casein Digest Medium			
Incubation Time	After 24h	After 48h	After 72h	After 96h	After120h	After144h	After168h
Observed Item	(—)	(—)	(—)	(—)	(—)	(—)	(—)
Positive Control	(+)						
Conclusion : Qualified Qualified							

Sterilized by EO,the products comply with ISO 11737 standard(Sterilization of medical devices--Microbiological methods--Part2:Tests of sterility performed in the validation of a sterilization process).

Note

(—) represent of negative BI,no bacterial growth

(+) represent of negative BI,bacterial growth

Drafted by Jane Zhang

Date :2021.9.8

Approved by Jane Zhang

Date : 2021.9.8

A red circular stamp with the text "EVERLASTING MEDICAL PRODUCTS" around the top and "BROWS" at the bottom. In the center, there is Chinese text "义乌市恒隆" and "义乌市恒隆". A handwritten signature in black ink, "Jane Zhang", is written across the stamp.