

**Foshan Flying Medical Products Co., Ltd**

**CERTIFICATE OF STERILIZATION ANALYSIS**

Manufacturer		Sleek Brows Ltd., for Everlasting Brows		Product Description		EVERLASTING BROWS HIGH QUALITY STAINLESS-STEEL MICROBLADING BLADES	
Country of Manufactory		P.R.China		Expiration Date		28/03/2026	
Size		20pcs/box		Lot Number		210329	
Item#		10CF0.18,10CF0.25,12CF0.18, 12CF0.25,14CF0.18,14CF0.25, 3*Shading		Inspection Date		21/03/2021	
PO#		20210306Z21		Customer Name		Everlasting Brows	
Flying PO#		20210306Z21		QTY		3250boxes	
NO	Test Item	Requirements	Inspection Standard	Inspected Qty.	Result	Conclusion	
1	Appearance	Clean, level off, no besmirch, no breakage	S-3	3250boxes	Accord	Pass	
2	Package Size	20pcs/box	S-2	3250boxes	Accord	Pass	
3	Box size	20mm*40mm*80mm			Accord	Pass	
4	Shear Strength	≥1N/cm			Accord	Pass	
5	Unwrap Quality	Good and ease to unwrap			Accord	Pass	
6	Seal	Sealed			Accord	Pass	
7	Printing Quality	Good quality good align				Pass	
8	Sterilization	Sterility			N/A	3250boxes	Accord
Final Conclusion			Inspection Passed				

**Drafted by Jane Zhang**  
**Date :2021.3.21**

**Approved by Jane Zhang**  
**Date :2021.3.21**

Foshan Flying Medical Products Co., Ltd

Sterilization Inspection Report

NO :

PO#	20210306Z21			Product	EVERLASTING BROWS HIGH QUALITY STAINLESS-STEEL MICROBLADING BLADES		
LOT	210329			Item#	10CF0.18,10CF0.25,12CF0.18,12CF0.25,14CF0.18,14CF0.25,3* Shading		
Sterilization Date	21/03/2021			Size	20pcs/box		
Sterilization Lot	210321005			Amount	3250boxes		
Process Parameter Summary				Inspection Basis	EN ISO 11737		
Temperature	50 ±2			Humidity	45-80%		
Exposure Time(h)	10h			EO Weight(kg)	61kgs		
<b>Biological Indicator Status</b>							
Manufacture	Hangzhou Future			Organism	ATCC 9372		
Lot	21031634			D-value	3.5min		
Population	2.2*10 C.F.U			Test Quantity	10+1 strips		
<b>Incubation Situation</b>							
Incubation Date	2021-3-21-2021-3-28			Incubation Temperature	35±2		
Incubation Time	168 小时			Medium Name	Soybean Casein Digest Medium 大豆-酪蛋白消化物培养基		
Incubation Time	After 24h	After 48h	After 72h	After 96h	After 120h	After 144h	After 168h
Observed Item	(-)	(-)	(-)	(-)	(-)	(-)	(-)
Positive Control	(+)						
Conclusion	: Qualified						
Qualified							

Sterilized by EO,the products comply with ISO 11737 standard(Sterilization of medical devices--  
Microbiological methods--Part2:Tests of sterility performed in the validation of a sterilization process).

Note

(-) represent of negative BI,no bacterial growth

(+) represent of negative BI,bacterial growth

Drafted by Jane Zhang

Date 2021.3.28

Approved by Jane Zhang

Date 2021.3.28

A red circular stamp with the text "EVERLASTING MEDICAL PRODUCTS" around the perimeter and Chinese characters in the center. Overlaid on the stamp is a handwritten signature in black ink that reads "Jane Zhang".